



**HAL**  
open science

## Multisensor fusion for atrial and ventricular activity detection in coronary care monitoring.

Alfredo Hernández, Guy Carrault, Fernando Mora, Laurent Thoraval,  
Gianfranco Passariello, Jean-Marc Schleich

► **To cite this version:**

Alfredo Hernández, Guy Carrault, Fernando Mora, Laurent Thoraval, Gianfranco Passariello, et al.. Multisensor fusion for atrial and ventricular activity detection in coronary care monitoring.. IEEE Transactions on Biomedical Engineering, 1999, 46 (10), pp.1186-90. 10.1109/10.790494 . inserm-00134401

**HAL Id: inserm-00134401**

**<https://inserm.hal.science/inserm-00134401>**

Submitted on 27 Mar 2007

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

# Multisensor Fusion for Atrial and Ventricular Activity Detection in Coronary Care Monitoring

Alfredo I. Hernández, Guy Carrault, Fernando Mora, Laurent Thoraval, Gianfranco Passariello and Jean M. Schleich

**Abstract**— Information management for critical care monitoring is still a very difficult task. Medical staff is often overwhelmed by the amount of data provided by the increased number of specific monitoring devices and instrumentation, and the lack of an effective automated system. Specifically, a basic task such as arrhythmia detection still produce an important amount of undesirable alarms, due in part to the mechanistic approach of current monitoring systems. In this work, multi-sensor and multi-source data fusion schemes to improve atrial and ventricular activity detection in critical care environments are presented. Applications of these schemes are quantitatively evaluated and compared with current methods, showing the potential advantages of data fusion techniques for event detection in noise corrupted signals.

**Index Terms**—CCU monitoring, P-wave detection, Ventricular activity detection, Systemic arterial pressure, Optimal decentralized data fusion.

## I. INTRODUCTION

Critical care monitoring has been the subject of many studies in order to increase its effectiveness, from the information management point of view. In this sense, it does not escape from the explosion of information sources and handling methodologies seen in other non-medical monitoring applications such as: strategic defense, satellite tracking, automatic supervision of industrial processes, petroleum pipeline transport surveillance, space flights, and so on. Proper integration of these sources and methodologies is a need in modern critical care monitoring.

Coronary Care Units (CCUs), have been dedicated to the follow up of cardiac patients under recovery, either during ischaemic episodes or after myocardial infarction. Mainly, they are devoted to the prevention and control of fatal and of non-fatal arrhythmias, and to specific therapeutic follow-up. In order to pursue this task, many physiological observations are routinely acquired in a typical CCU, among them: several channels of ECG, blood pressure, blood chemistry, clinical observations, and even echo doppler data. These sensors and data are not easily combined by machines, mainly due to the mechanistic approach in which they are implemented, and the differences in data sources. The information "fusion", to produce a meaningful interpretation for the decision making process, is left to the experience of the attending cardiologist and nurses whose responsibility goes as far to collect, classify, organize and

interpret massive data, both redundant and complementary, in order to construct an appropriate mental image of the pathophysiological state of the patient [1].

Attempts to reduce this information handling stress in time-critical environments have been undertaken by many researchers and groups. Automatic arrhythmia detection has been the subject of extensive work, due to the need of a fast response when life-threatening events occur. Nevertheless, most systems still produce undesirable alarms and are prone to confuse artifacts with rhythms, becoming quite ineffective in every day practice, and making human observation and monitoring mandatory. Currently, mostly ventricular arrhythmias are detected. Ischaemic episode trending requires special methodologies which are still under development [2] and atrial arrhythmias have proven to be a rather elusive problem due to the difficulties inherent to *P*-wave detection.

This paper presents some ideas to enhance current methods of cardiac rhythm monitoring. The approach undertaken seeks to integrate, with the usual ECG signals, complementary data from haemodynamic processes or from other electrical signals such as the esophageal ECG (EECG). A data fusion paradigm [3] is proposed to handle the information integration approach. As such, the following aspects have been developed: *i*) Data fusion examples, combining multi-sensor and multi-source information *ii*) Presentation of explicit technical solutions and, *iii*) Quantitative performance measurements for the particular tasks chosen, by means of Receiver Operational Characteristics (ROC) curve analysis [4].

In this work, data fusion techniques have been applied to the detection of fundamental cardiac events, namely the ventricular activity (VA) and the atrial activity (AA) –represented respectively by the *QRS* complex and the *P*-wave in the ECG. Evaluation of AA and VA detection schemes has been performed for an initial set of real and simulated signals, in order to present the advantages of the data fusion approach. With a proper detection of these two activities, a more accurate characterization of the cardiac rhythm can be achieved, particularly for differential diagnoses between ventricular and supra-ventricular tachycardia [5].

## II. CLINICAL EXAMPLES OF DATA FUSION

### *A. Association of ECG and haemodynamic signals*

Haemodynamic signals are analyzed usually in a punctual basis either during diagnostic procedures in the cath lab, or during continuous bedside monitoring. Although being of heterogeneous nature, both the ECG and haemodynamic signals, such as the Impedance Cardiograph or Blood Pressure, have information mutually correlated due to the physiological interrelation of the mechanical and electrical functions of the heart. In medical practice, this existing interconnection has been very rarely exploited for rhythm analysis purposes.

A simple example of this effect is shown in figure 1. The ECG signal by itself, even using three leads, would be interpreted by an automatic recognition system as showing ventricular couplets. Further analysis of the haemodynamic signal invalidates this hypothesis, suggesting that the second event can be interpreted as an artifact by simply observing the absence of a systolic peak in concordance with the second candidate QRS, since noise sources in the ECG and the haemodynamic signal are decorrelated.

In critical situations where the ECG is missing (leads off), or degraded (patient movement or electrosurgery) classic detection systems behave like an erratic alarm generator. Joint analysis by means of additional sensors, such as the ventricular pressure signal, may allow the preservation of a window to monitor the cardiac rhythm, assuring patient monitoring. Nowadays, data fusion with electromechanic information still comprises several obstacles to its application in automatic supervision, mainly because they are obtained by invasive techniques. Nevertheless, these procedures are still widely used in current CCU units, awaiting newer non invasive ways to collect information from haemodynamic processes, such as automatic oscillometric blood pressure measurement.

INSERT FIGURE 1 HERE

Fig.1. Multi-source fusion (electrical and mechanical): 3 ECG leads (upper panel) and intra-ventricular pressure (lower panel).

Analysis of ECG traces suggest the presence of two ventricular couplets. Analysis of pressure curves allows us to invalidate this hypothesis and to re-establish the more probable diagnosis of a unique ventricular extrasystole followed by an artifact. The presence of a premature ventricular contraction (PVC) in the ECG lead induced a ventricular pressure fall linked to the short coupling time (point 1) and an increase of pressure superior than the average (Point 2) due to the compensation rest following a PVC (the ventricle presents a lengthened filling time).

### *B. ECG-EECG association*

The interest in this case is the differential diagnosis between supra-ventricular and ventricular arrhythmias, observing the heart's electrical function alone. In addition to the surface ECG, the EECG is used to obtain an enhanced atrial activity representation (*A*-wave) which directly correlates to the *P*-wave in the surface ECG. Some proposed methods of automatic arrhythmia recognition perform essentially the location and identification of the *A*-wave, allowing the analysis of the atrio-ventricular conduction and its dysfunction [6,7]. Enhancement of the atrial electrical activity in the EECG allows to distinguish supra-ventricular arrhythmias with aberrant conduction from arrhythmias of ventricular origin, such as those depicted figure 2. This task is still rather difficult to perform in current CCU applications, due to the possibility of wave superposition (*i.e.* *P*-wave immersed in the *QRS* or *T*-waves). Besides its application on supra-ventricular arrhythmias, fusion of the surface ECG and the EECG allows the distinction between sinus tachycardia,

atrial flutter, tachysystole, atrial fibrillation, and also between tachycardia with intra-nodal reentries and those using a secondary path (beam of Kent)[8,9].

INSERT FIGURE 2 HERE

Fig. 2. Example of electrical sources printouts (a) ECG (upper panel); (b) Acquired esophageal ECG (EECG) (middle panel); and (c) filtered EECG (lower panel). Although the  $P$ -wave is not visible on the whole surface ECG printout, the atrial activation wave in the EECG ( $A$ -wave) is particularly predominant, notably after signal processing of the EECG. Analysis of the EECG allows us to refute the hypothesis of ventricular tachycardia that could have been diagnosed by the analysis of the ECG alone (a) and keep the more probable diagnosis of supraventricular tachycardia with bundle branch block.

### III. MULTISENSOR FUSION APPROACHES

A general structure applicable to both VA and AA detection is presented in figure 3. It is based in a distributed detection scheme of  $n \geq 2$  pre-processors fed by a set of  $k \geq 1$  different sensors ( $S_{k,n}$ ). Each pre-processor is composed of: *i*) a nonlinear transformation to enhance the activity of interest, *ii*) a sensor association stage (SA) in order to minimize false alarms, and *iii*) a global fusion node of local decisions. Every pre-processor produces a local detection of the activity of interest ( $u_1, \dots, u_n$ ) which is merged by a general data fusion stage, to produce a final decision  $u$  by means of a fusion rule  $f(u_1, \dots, u_n)$ .

Two realizations of this general structure have been implemented and evaluated: the first one is focused on VA detection based on ECG and pressure signals, and the second on AA detection from the ECG and the EECG. In each case, particular pre-processors have to be developed. They are described in the following sections.

INSERT FIGURE 3 HERE

Fig. 3. General detector structure based on a distributed detection scheme.

#### A. Ventricular activity detection

In this section, mono-source and multi-source VA detection methods are presented. In the case of mono-source detection, two pre-processors are used, each one fed by a different ECG lead ( $n=2$ ,  $S_{1,1}$ =ECG lead I,  $S_{1,2}$ =ECG lead II). For the multi-source implementation, three pre-processors are used: two pre-processors are fed by different ECG leads, as previously, and the third by a SAP signal ( $n=3$ ,  $S_{1,1}$ =ECG lead I,  $S_{1,2}$ =ECG lead II,  $S_{1,3}$ =SAP).

The first level in each pre-processor acts as a pre-detector of ventricular activity. It is based on the use of the following methods:

- *QRS* detector reported by Gritzali [10],
- A matched filter, adapted to the systolic ejection slope of the haemodynamic signal, for the detection of the increase of ventricular pressure [11].

In low-noise situations, the same number ( $N_v$ ) of ventricular events are estimated from the *QRS* and systolic ejection detectors, with a time delay between them. For a given patient showing a regular sinus rhythm, this physiological delay ( $\hat{\delta}$ ) between the electrical and mechanical myocardial activation can be considered as a random variable following a Normal law  $P_A(\hat{\delta})$  with mean  $\mu_{EMD}$  and variance  $\sigma_{EMD}$ , as suggested by Thoraval in [12]. Also, the association stage appears quite simple and consists to artificially re-synchronize the electrical and mechanical activities by introducing a delay of  $\hat{\delta}=\mu_{EMD}$  on the ECG channels, in order to assure the coherence of decisions taken on each pre-processor, before the fusion node.

### B. Atrial activity detection

Two pre-processors are used in this implementation ( $n=2$ ). Each one has two input signals: one providing mainly a good representation of VA, ( $S_{1,i}=S_{V,i}$ ) and the second ( $S_{2,i}=S_{A+V,i}$ ) being a good index of AA. Obviously, this second signal will also present ventricular information, which will be considered as impulsive noise for that channel. Pre-processor 1 will acknowledge a decision when a *P*-wave is detected in the ECG ( $S_{A+V,1}=S_{V,1}=ECG$ ), and Pre-processor 2 will acknowledge a decision when an *A*-wave is detected in the EECG ( $S_{A+V,2}=EECG$ ;  $S_{V,2}=ECG$ ).

INSERT FIGURE 4 HERE

Fig. 4. Particular pre-processor implementation for AA detection.

Actual preprocessor implementation for AA detection is done as follows [13]: The first level is based on the method proposed by Thakor and Yi-Sheng [14] which cancels the ventricular activity by means of the *QRS-T* adaptive filter  $AF_1$ . This adaptive filter is fed by VA detection impulses ( $i_l$ ), previously generated by modules  $D_l$  and  $T_{si}$ . More precisely,  $AF_1$  is an impulse-response filter having as weights an inverted template of the *QRS-T*, adaptively calculated. Gritzali's *QRS* detection algorithm [10] was used for  $D_l$ .

The second level performs the sensor association and involves a second adaptive filter  $AF_2$  (*i.e.* optimal Wiener filtering). Module  $D_2$  consists of a low-pass filter ( $F_l$ ) tuned to the spectral band of the *P*-wave, and a square-law envelope transformation. The resulting signal ( $S_2$ ) is introduced in  $AF_2$ , in which an index of the undesired ventricular

activity ( $S_1$ ) is used as reference. This level aims to eliminate remaining ventricular activity in  $S_2$ , mainly due to the detection jitter, sharp morphological changes (such those produced by a PVC) or missed  $QRS$  detections, by taking into account the correlated information between  $S_1$  and  $S_2$ .

The third level proceeds to the detection of  $P$ -waves by means of a classical filter ( $F_2$ ) and thresholding, to obtain a local atrial activity detection ( $u_i$ ), used as input to the final data fusion stage. Design of filters  $F_1$  and  $F_2$  is based on the spectral analysis of several  $P$ -waves and  $QRS$  complex residuals, in order to separate at best the two spectral bands.

#### IV. GENERAL DATA FUSION

From a decisional point of view, each pre-processor has as objective to test the two following hypotheses:

$$\begin{aligned} H_0 : x_i(t) &= n_i(t) + A_i(t) \\ H_1 : x_i(t) &= n_i(t) + A_i(t) + W_i(t - \tau_{ij}) \end{aligned}$$

where  $W_i(t - \tau_{ij})$  is the  $j^{\text{th}}$  wave  $W$  observed on sensor  $i$ , appearing at instant  $\tau_{ij}$ , and  $n_i(t)$  and  $A_i(t)$  are stationary and impulsive noises respectively.  $W_i(t - \tau_{ij})$  corresponds to, depending on the event to be detected, the  $P$ -wave (ECG) or  $A$ -wave (EECG) for AA detection, or the  $QRS$  complex (ECG) and the rising edge of the ventricular pressure wave in the case of VA. Each pre-processor should produce therefore an independent boolean output such that, under the hypothesis

$$H_0: u_i = -1 \text{ and under the hypothesis } H_1: u_i = 1.$$

The fusion problem consists on producing a new statistical  $u = f(u_1, \dots, u_n)$  which represents a local decision association. The chosen function  $f$  is a sum of local decisions, weighted by the estimated performance of each pre-processor's output (known as optimal fusion) determined by Chair and Varshney [15].

$$f(u_1, \dots, u_n) = \begin{cases} 1, & \text{if } a_0 + \sum_{i=1}^n a_i u_i > 0 \\ -1, & \text{otherwise.} \end{cases}$$

where optimum weights  $a_i$  are a function of the probability of false alarm  $P_{FAi}$  and miss  $P_{Mi}$  for sensor  $i$ , and the  $a$  priori probabilities of each hypothesis ( $P_0 = P(H_0); P_1 = P(H_1)$ )

$$\begin{aligned} a_0 &= \log \frac{P_1}{P_0} \\ a_i &= \log \frac{1 - P_{Mi}}{P_{FAi}} \quad \text{if } u_i = +1 \\ a_i &= \log \frac{1 - P_{FAi}}{P_{Mi}} \quad \text{if } u_i = -1. \end{aligned}$$

These weights represent the reliability of decision  $u_i$  based on the operational characteristics of detector  $i$ .

For all performance measurements, sensor fusion was calculated for equivalent  $P_{FA}$  on each local detector involved. Probabilities  $P_0$  and  $P_1$  are estimated from the annotations of the event of interest.

## V. RESULTS

### A. Detection of ventricular activity

Before proceeding further, some important aspects deserve to be clarified. In spite of the fact that multi-sensor detection of the *QRS* complex is supposed to be a relatively well known problem nowadays, nevertheless, very few works have attempted to describe the mathematical laws which govern the optimal or the most suitable association scheme for accurate event detection. Regarding this point, several questions can be posed. Do the association have to be global as suggested by Gritzali? Or can it make use of local detections with a fusion of decisions, in a decentralized detection scheme?

Results of multi-source fusion were obtained using the IMPROVE European database, corresponding to recordings collected during 24-hour ICU monitoring [12]. This database is constituted of multiple signals from different sources, including two ECG channels, systemic arterial pressure (SAP), pulmonary arterial pressure (PAP), central venous pressure (CVP), carbon dioxide concentration (CO<sub>2</sub>), oxygen concentration (O<sub>2</sub>), air flow (AWF) and the airway pressure (AWP). Patient record KUO0007 was used for testing. This particular recording was chosen because it presents many periods of high and low frequency noise and sensor failure, as well as different pathological rhythms. Several segments were selected from the record and concatenated together in order to artificially reconstruct a recording with 5738 cardiac cycles, containing intermixed noisy and clean periods. Also, annotations were made for each ventricular activity representation in the ECG and SAP channels.

Figure 5 presents ROC curves obtained from each individual sensor (ECG1, ECG2 and SAP), from a centralized detection scheme based on the multidimensional length transformation proposed by Gritzali (GRIT2D), and from distributed detection with optimal data fusion of the three available sensors (OPT). Optimal data fusion was also calculated for ECG sensors alone showing the same curve of ECG lead 1 -due to the reduced performance of the local detector applied to ECG lead 2-, and providing higher performances than the centralized detection (GRIT2D).

INSERT FIGURE 5 HERE

Fig. 5. ROC curve of ventricular activity detection measured from record KUO007 of the IMPROVE European database.

Also, a previous study on 10 files of the MIT-BIH arrhythmia database [16], containing two ECG channels, has been undertaken in our laboratory. Applying Gritzali's algorithm for each individual sensor (ECG1 and ECG2), it emerged from this study that the decentralized approach with an optimal local fusion (OPT) presents superior performances than those shown by the centralized approach (GRIT2D). These results are consistent with those shown in figure 5.

Finally, regarding the multi-source data fusion, results show that the gain in detection accuracy obtained by jointly considering both electrical and mechanical sources produced by the cardiovascular system is quite important when compared to the mono-sensor analysis of the ECG signal.

### B. Detection of atrial activity

Validation of these methods was undertaken using real signals recorded in an ambulatory setting at the “Fundación Venezolana de Cardiología”. The selected record is composed of a single ECG lead and an esophageal ECG channel.

The proposed structure was compared with three methods previously presented in the literature for AA detection: those by Gritzali [17], Thakor [14] (to which a supplementary detection stage was added), and Jenkins[6]. For the EECG, Jenkins proposed a simple scheme for *A*-wave detection based on filtering and thresholding. In this work, Gritzali's algorithm, included into the preceding detection scheme, was adapted to the case of the *A*-wave in the EECG signal.

1) *Sensor Association*: Performance evaluation at the second level of each pre-processor (which consists to reduce the impulsive noise related to *QRS* suppression) was done in order to quantify the gain obtained in this stage. An estimation of the signal-to-noise ratio (SNR) at the output of each level is presented in this section for the case of pre-processor 2 (combination of EECG and ECG signals). The SNR for beat *b* and sensor *s*, is defined as:

$$SNR_b^s = 10 \log_{10} \left( \frac{\hat{E}_b(P^2)}{\hat{E}_b(QRS^2)} \right)$$

where  $\hat{E}(P^2)$  and  $\hat{E}(QRS^2)$  represent an estimate of the energy of the *P*-wave and *QRS* complex respectively. A boxplot is presented in figure 6 showing the increase in the SNR accomplished by the association stage, and clearly justifying its utilization.

INSERT FIGURE 6 HERE

Fig. 6. Boxplot of the SNR between the energies of the atrial and ventricular activities from a surface ECG, an EECG and the output of level 1 (Prep2, L1) and level 2 (Prep2, L2) for pre-processor 2.

2) *Detection Performance*: Figure 7 shows performance measurements for AA detection. From these results, the next points can be emphasized: *i*) The detector proposed by Gritzali [17] do not constitute a satisfactory solution; *ii*) Isolated analysis of a single EECG lead does not present a valid solution -this technique remains sensitive to frequent movements of the sensor, producing significant morphological changes of *A* and *V* waves, therefore minimizing the SNR ; *iii*) Performance measurement at the output of the first level of pre-processor 1 (i.e. the method proposed by Thakor [14]) are interesting, but are improved by the adaptive filter that associates ECG and EECG signals. A higher gain in performance, due to the association stage in level 2, is shown by the pre-processor 2 (see also figure 6).

INSERT FIGURE 7 HERE

Fig. 7. ROC curve of atrial activity detection measured from ECG and EECG signals.

Overall detector performance is improved even more at the output of the global fusion stage, showing a high detection probability ( $P_d > 0,8$ ) even for very low values of false alarms ( $P_{FA} < 10^{-4}$ ). These results also show how a local detector that does not present *a priori* interesting performance characteristics (i.e. output of pre-processor 1), can be exploited to enhance the probability of detection of the overall scheme, for a given probability of false alarm. It is also important to note the increased performance of pre-processor 1, compared to classical methods reported in the literature.

3) *Robustness Evaluation*: Performance of the proposed detector in face of: *i*) atrio-ventricular dissociation, *ii*) retrograde *P*-waves, *iii*) Extrasystoles (morphological variations of the *QRS* complex), *iv*) Miss-detections of the *QRS* complex and *v*) jitter on the *QRS* detection stage, has been measured using simulated signals. It is important to notice that all of these parameters can notably affect detection performance.

INSERT FIGURE 8 HERE

Fig. 8. ROC curve of the sensibility of pre-processor 2 (EECG and ECG) to atrio-ventricular block; Pre-processor 2 is somewhat insensitive (i.e. given a fixed  $P_{FA}$ , the probability of detection  $P_d$  does not fluctuate) to AV dissociation.

Figure 8 presents ROC curves of pre-processor 2, calculated from simulated signals for different probabilities  $p$  of appearance of a dissociated AA. For the totality of the tests (not described here due to the limited space) the proposed solution always presents superior results than the other techniques, and exhibits on the other hand, to a fixed rate of false alarms, a lower relative deviation of detection probability.

## VI. CONCLUSION

Two schemes, based on a proposed general structure for multi-sensor event detection, were implemented and tested in this work: one for AA detection and another for VA detection. Three different preprocessor implementations have been used in the evaluated schemes, for sensor association and local detection, depending on the types of signals used as inputs, and the activity to be detected. The interest of this work was not to substitute the clinical use of the surface ECG, which still remains as far the more robust technique to characterize cardiac arrhythmias, but to simply show that automatic monitoring of the cardiovascular system can be vastly improved by combining the ECG with other signals.

Results show clearly the improvement in performance and robustness obtained from both the sensor association stage and the final atrial and ventricular activity detection, when different types of cardiovascular signals are combined into a data fusion model, suggesting that the presented detection schemes can lead to a better automatic classification of cardiac arrhythmias.

Regarding VA detection, the multi-source scheme showed a greater  $P_d$  for a given  $P_{FA}$  than any other evaluated method. A fixed delay was used to associate asynchronous events detected from different sources. Also other association methods, such as adaptive estimation of time delays, are currently being evaluated. In the multi-sensor detection of AA, based on surface and esophageal ECG leads, the proposed detection structure provided an increased detection performance and robustness.

Obviously, the amount of redundant and complementary information obtained from each local detector is crucial to the fusion process. In most cases, sensors observing different sources of the same activity of interest are more likely to offer complementary information and to produce increased detection performances. However, important gains in performance are also obtained from multi-sensor/mono-source detection schemes, by means of the decentralized optimal rule.

It is important to notice that the implementation of the fusion rule used in this work depends on an on-line or batch estimation of the performance characteristics of each local detector ( $P_{FAi}$ ,  $P_{Mi}$ ), and *a-priori* probabilities of the confronted hypothesis ( $P_0$ ,  $P_1$ ). Development in this sense is being carried out as part of our project in intelligent monitoring, concerning the adaptation of the proposed algorithms for real-time applications and real-time algorithm switching, based on the estimation of the acquisition environment state (clean, noisy, very noisy) [18].

Although the proposed data fusion model has been only applied in this work to the UCC/ICU environment, its extension to other biomedical applications such as pacemaker control and intra-cardiac defibrillation is also possible. In the case of pacemaker control, an esophageal signal could acknowledge the correct use and placement of pacing electrodes, and the stability of the P-wave detected by the pacemaker [19]. In the case of intra-cardiac defibrillation, the

jointed evaluation of electrical stimuli and its mechanical response could lead to optimize the quality and energy of the stimuli.

#### ACKNOWLEDGEMENTS

The authors would like to thank Dr. María I. Hernández, and the Fundación Venezolana de Cardiología, for the acquisition and annotation of ECG and EECG signals.

This work has been supported in part by Project BID-CONICIT E-08, and the PCP-GBM Cooperation Program.

#### REFERENCES

- [1] F. Mora, G. Passariello, G. Carrault, J. P. Le Pichon, "Intelligent patient monitoring and Management Systems : A review," *IEEE Eng Med, Biol. Mag.*, vol. 12, pp. 23-33, 1993.
- [2] A. Bosnjak, G. Bevilacqua, F. Mora F, G. Passariello, B. Sanso, G. Carrault, "An approach to intelligent Ischaemia monitoring," *Med. Biol. Eng. Comp.*, vol. 33, pp. 749-756, 1995.
- [3] D. Hall, *Mathematical techniques in multisensor data fusion*. Artech House, MA, 1992.
- [4] J. Michaelis, J. Willems, "Performance Evaluation of Diagnostic ECG Programs" *In : Computers in Cardiology 1988*. Los Alamitos: IEEE Computer Society Press, pp. 25-30, 1988.
- [5] R. LeBlanc, "Quantitative Analysis of Cardiac Arrhythmias," *CRC Crit. Rev. in Biom.. Eng.*, vol. 14, no.1, pp. 1-43, 1986.
- [6] J. Jenkins, "Automated Electrocardiography and Arrhythmia Monitoring," *Prog. Card. Dis.* vol. 25. no. 5, pp. 367-408, 1983.
- [7] J. Jenkins, R. C. Arzbaecher, "Computer diagnosis of Supraventricular and Ventricular Arrhythmias. A new esophageal technique," *Circulation*, vol. 60, pp. 977-987, 1979.
- [8] W. H. Brown, "A study of the esophageal lead in clinical electrocardiography, Part I," *Am. Heart J.*, vol. 12, pp. 1-45, 1936.
- [9] W. H. Brown, "A study of the esophageal lead in clinical electrocardiography, Part II," *Am. Heart J.*, vol. 12, pp. 307-381, 1936.
- [10] F. Gritzali, "Towards a generalized scheme for QRS detection in ECG waveforms," *Signal Processing*, vol. 15, pp. 183-192, 1988.
- [11] I. Pino, *Aplicación de la Fusión de Datos para la detección robusta de eventos cardíacos*. Masters Thesis. Caracas, Venezuela, Universidad Simón Bolívar, 1998.
- [12] "The IMPROVE Project," *IEEE Eng Med, Biol. Mag.*, vol. 16, no. 6, 1997.

- [13]A.I. Hernández, G. Carrault, F. Mora, G. Passariello, M.I. Hernández, J Schleich. "Quantitative Comparison of Atrial Activity Detection Approaches". In: *Computers in Cardiology 1997*. Los Alamitos: IEEE Computer Society Press, pp. 481-484, 1997
- [14]N. Thakor, Z. Yi-Sheng. "Applications of Adaptive Filtering to ECG Analysis: Noise Cancellation and Arrhythmia Detection," *IEEE Trans. Biom. Eng*, vol. 38, no. 8, pp. 785-794, 1991.
- [15]Z. Chair, P. Varshney. "Optimal data fusion in multiple sensor detection systems," *IEEE Trans. Aer. Electr. Syst*, vol. AES-22, no. 1, pp. 98-101, 1986.
- [16]G. B. Moody, R. G. Mark, "The MIT-BIH Arrhythmia Database on CD-ROM and Software for use with it," In : *Computers in Cardiology 1990*. Los Alamitos: IEEE Computer Society Press, pp. 185-188, 1990.
- [17]F. Gritzali, G. Frangakis, G. Papakonstantinou, "Detection of the P and T waves in an ECG," *Comp. Biom. Res.*, vol. 22, pp. 83-91, 1989.
- [18]G. Le Certen, T. Soulas, G. Carrault, J. P. Le Pichon,"SIMBAD: Intelligent ECG Monitoring System (A real-time software implementation)," *Proc. 8<sup>th</sup> Int. IMEKO Conf. on Meas. in Clin. Med.*, pp. 3.11-3.14, 1998.
- [19]M. G. Bongiorni, E. Soldati, L. Paperini, "Evaluation of rate-responsive pacemakers by transesophageal Holter monitoring of spontaneous atrial rate," *PACE*, vol. 13, pp. 1755-1760, 1990.




Fig.1. Multi-source fusion (electrical and mechanical): 3 ECG leads (upper panel) and intra-ventricular pressure (lower panel). Analysis of ECG traces suggest the presence of two ventricular couplets. Analysis of pressure curves allows us to invalidate this hypothesis and to re-establish the more probable diagnosis of a unique ventricular extrasystole followed by an artifact. The presence of a premature ventricular contraction (PVC) in the ECG lead induced a ventricular pressure fall linked to the short coupling time (point 1) and an increase of pressure superior than the average (Point 2) due to the compensation rest following a PVC (the ventricle presents a lengthened filling time).




Fig. 2. Example of electrical sources printouts (a) ECG (upper panel); (b) Acquired esophageal ECG (EECG) (middle panel); and (c) filtered EECG (lower panel). Although the *P*-wave is not visible on the whole surface ECG printout, the atrial activation wave in the EECG (*A*-wave) is particularly predominant, notably after signal processing of the EECG. Analysis of the EECG allows us to refute the hypothesis of ventricular tachycardia that could have been diagnosed by the analysis of the ECG alone (a) and keep the more probable diagnosis of supraventricular tachycardia with bundle branch block.




Fig. 3. General detector structure based on a distributed detection scheme.




Fig. 4. Particular pre-processor implementation for AA detection.




Fig. 5. ROC curve of ventricular activity detection measured from record Kuo007 of the IMPROVE European database.




Fig. 6. Boxplot of the SNR between the energies of the atrial and ventricular activities from a surface ECG, an EECG and the output of level 1 (Prep2, L1) and level 2 (Prep2, L2) for pre-processor 2.




Fig. 7. ROC curve of atrial activity detection measured from ECG and EECG signals.




Fig. 8. ROC curve of the sensibility of pre-processor 2 (EECG and ECG) to atrio-ventricular block; Pre-processor 2 is somewhat insensitive (*i.e.* given a fixed  $P_{FA}$ , the probability of detection  $P_d$  does not fluctuate) to AV dissociation.