



HAL
open science

First attempt to motion corrected flow encoding using free-breathing phase-contrast CINE MRI

Christophe Meyer, Pierre-Andre Vuissoz, Damien Mandry, Jacques Felblinger

► To cite this version:

Christophe Meyer, Pierre-Andre Vuissoz, Damien Mandry, Jacques Felblinger. First attempt to motion corrected flow encoding using free-breathing phase-contrast CINE MRI. 15th Annual SCMR Scientific Sessions, Feb 2012, Orlando, United States. pp.W53, 10.1186/1532-429X-14-S1-W53 . inserm-00665426

HAL Id: inserm-00665426

<https://inserm.hal.science/inserm-00665426v1>

Submitted on 1 Feb 2012

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

WORKSHOP PRESENTATION

Open Access

First attempt to motion corrected flow encoding using free-breathing phase-contrast CINE MRI

Christophe Meyer^{1,3*}, Pierre-Andre Vuissoz^{1,3}, Damien Mandry^{2,3}, Jacques Felblinger^{1,4}

From 15th Annual SCMR Scientific Sessions
Orlando, FL, USA. 2-5 February 2012

Summary

This study demonstrates the feasibility of free-breathing phase-contrast CINE MRI without averaging. A new version of the CINE GRICS algorithm[1] was used to correct for motion.

Background


Phase-contrast MRI encodes speed and direction of moving spins by means of toggling a bipolar gradient. It is a valuable tool for assessing conditions affecting the vascular system by measuring the velocity of flowing blood[2]. Clinically, this sequence is performed in breath-hold or in free breathing but, in the latter case, using signal averaging. We propose to demonstrate the feasibility of free-breathing phase-contrast CINE MRI without averaging exploiting the acquisition redundancy by applying a new version of the CINE GRICS algorithm[1] to correct for motion.

Methods

Cardiac examination (approved by our local ethics committee) was performed on one normal volunteer during which three 2D phase-contrast CINE MRI sequences (common parameters: 256x128 acquisition matrix, 6 views per segment (vps), 32 reconstructed cardiac phases, 150 cm/sec VENC, slice direction velocity encoding, 5 mm slice thickness, 44 cm FOV, 62.5 kHz bandwidth, 3.05/8.08 ms TE/TR, 15° flip angle) were acquired on a 3T scanner (Signa HDxt, GE Healthcare, Milwaukee, WI) with a 8-element cardiac coil : (1) breath-held (2) averaged (3 NEX) in free breathing (3) in free breathing storing the raw data of 3 NEX to an external computer for offline processing. Signals from a respiratory belt were carried by a custom Maglife patient

monitoring system (Schiller Medical, France) and recorded with a dedicated home-made hardware.

Offline processing consisted of splitting the raw data from the 2 velocity encoding steps thus giving 2 sets of



¹IADI / INSERM U947, Vandoeuvre-les-Nancy, France
Full list of author information is available at the end of the article




Figure 2 Velocity curves in an ascending aorta ROI (D) from phase contrast CINE scans in breath-hold (A) in free breathing with averaging (C) and in free breathing with PC CINE GRICS (B).

raw data. They were formatted and processed separately, along with the respiratory signals, by the CINE GRICS algorithm on a 16-node compute cluster. Total reconstruction delay was 15 min. Phase difference images (Figure 1) were created. Images were analysed by an experimented radiologist using CV Flow (GE Medical Systems) and velocity curves were produced (Figure 2) for the ascending aorta ROI.

Results

Figure 1 shows a section of the aortic arch from the 3 sequences (all same plane coordinates and cardiac phase). Averaged acquisition exhibits blurring on the images generated by the manufacturer's software. Motion correction using GRICS removes almost all ghosting artefacts and improves vessel contrast delineation. Figure 2 shows that velocity curves from the breath-held, free breathing and PC CINE GRICS scans present the same features (note that curves are shifted due to trigger delay implementations).

Conclusions

We have demonstrated that motion corrected free-breathing phase-contrast CINE MRI using GRICS is

feasible. Future work will focus on (1) a 3D PC CINE implementation and (2) increased resolution for sharp, small structures imaging.

Author details

¹IADI / INSERM U947, Vandoeuvre-les-Nancy, France. ²CHU Nancy, Vandoeuvre-les-Nancy, France. ³Université de Lorraine, Nancy, France. ⁴CIC-IT Nancy, Nancy, France.

Published: 1 February 2012

References

1. Vuissoz, et al: Free-Breathing Imaging of the Heart Using 2D Cine-GRICS (Generalized Reconstruction by Inversion of Coupled Systems) With Assessment of Ventricular Volumes and Function. *J. of Magn Reson Imag.*
2. Rebergen, et al: *Am. Heart J.* 1993.

doi:10.1186/1532-429X-14-S1-W53

Cite this article as: Meyer et al.: First attempt to motion corrected flow encoding using free-breathing phase-contrast CINE MRI. *Journal of Cardiovascular Magnetic Resonance* 2012 **14**(Suppl 1):W53.