



Intra-and inter-rater reproducibility of ultrasound imaging of patellar and quadriceps tendons in critically ill patients

Joana Castro, Karina Livino de Carvalho, Paulo Eugênio Silva, Emerson Fachin-Martins, Nicolas Babault, Rita de Cássia Marqueti, João Luiz Quagliotti Durigan

► To cite this version:

Joana Castro, Karina Livino de Carvalho, Paulo Eugênio Silva, Emerson Fachin-Martins, Nicolas Babault, et al.. Intra-and inter-rater reproducibility of ultrasound imaging of patellar and quadriceps tendons in critically ill patients. PLoS ONE, 2019, 14 (6), pp.e0219057. 10.1371/journal.pone.0219057 . inserm-02183508

HAL Id: inserm-02183508

<https://inserm.hal.science/inserm-02183508>

Submitted on 15 Jul 2019

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

RESEARCH ARTICLE

Intra- and inter-rater reproducibility of ultrasound imaging of patellar and quadriceps tendons in critically ill patients

Joana Castro^{1,2☯*}, Karina Livino de Carvalho^{1☯}, Paulo Eugênio Silva^{2,3☯}, Emerson Fachin-Martins^{1,3☯‡}, Nicolas Babault^{4‡}, Rita de Cássia Marqueti^{1,3☯‡}, João Luiz Quagliotti Durigan^{1,3‡}


1 Graduate Program in Rehabilitation Sciences, University of Brasília (UnB), Ceilândia, Federal District, Brazil, **2** Institute of Strategic Health Management of the Federal District (IGESDF), Brasília, Brazil, **3** Graduate Program in Science and Technology in Health, University of Brasília (UnB), Ceilândia, Federal District, Brazil, **4** INSERM UMR1093-CAPS, Université Bourgogne Franche-Comté, UFR des Sciences du Sport, Dijon, France

☯ These authors contributed equally to this work.

‡ Current address: Physiotherapy Department, Ceilândia College, University of Brasília (UnB), Ceilândia, Federal District, Brazil

‡ These authors also contributed equally to this work.

* joanadtcastro@gmail.com



OPEN ACCESS

Citation: Castro J, Livino de Carvalho K, Silva PE, Fachin-Martins E, Babault N, Marqueti RdC, et al. (2019) Intra- and inter-rater reproducibility of ultrasound imaging of patellar and quadriceps tendons in critically ill patients. PLoS ONE 14(6): e0219057. <https://doi.org/10.1371/journal.pone.0219057>

Editor: Shane Patman, University of Notre Dame Australia, AUSTRALIA

Received: October 16, 2018

Accepted: June 15, 2019

Published: June 27, 2019

Copyright: © 2019 Castro et al. This is an open access article distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Data Availability Statement: All relevant data are within the manuscript.

Funding: This work was supported by the Fundação de Amparo à Pesquisa do Distrito Federal (FAPDF; Process number: 193.000.862/2014); Conselho Nacional de Desenvolvimento Científico e Tecnológico (CNPq Process number: 447529/2014-5; 310359/2014-7) and by CAPES (Programa Ciências sem Fronteiras, Pesquisador Visitante Especial; Process number: 403

Abstract

Since the outset of body image reconstruction for diagnosis purposes, ultrasound has been used to investigate structural changes located in tendons. Ultrasound has clinical applications in the intensive care unit, but its utility for tendon imaging remains unknown. Thus, we aimed to determine intra- and inter-rater reproducibility of measures obtained by images generated through morphological tendon sonographic analysis recorded from critically ill patients. We designed a cross-sectional study to assess thickness, cross-sectional area, and echogenicity of patellar and quadriceps tendons in a convenience sample formed with 20 critically ill patients. Two independent raters (experienced and novice) recorded repeated measures, checking for agreement (Kappa statistics) and reliability (Intraclass coefficient Correlation-ICC and Bland-Altman). The quality of images acquired by the two independent raters substantially agreed ($k = 0.571-1.000$), regardless of the region on the patellar tendon or the studied tendon (patellar or quadriceps). Regardless of how much experience the rater had, their repeated records (intra-rater reliability) always demonstrated almost complete correlation, ICC ranging from 0.89 to 0.98 for both tendons in all outcomes. At the same way, the statistically significant inter-rater ICC ranging from 0.87 to 0.97. Both repeated measures by the raters (intra-rater) and the repeated single and double measures between the raters (inter-rater) presented a minimum measurement error constituting a predominant pattern of random variability. We conclude that ultrasound imaging acquisition performed by independent raters for tendon thickness, CSA, and echogenicity monitoring of critically ill patients are acceptable and are not influenced by rater experience.

88881.068106/2014-01). This study was financed in part by the Coordenação de Aperfeiçoamento de Pessoal de Nível Superior – Brasil (CAPES) – Finance Code 001. The funders had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript.

Competing interests: The authors have declared that no competing interests exist.

Introduction

Critically ill patients suffer from extensive muscle wasting and atrophy, which occurs rapidly at the onset of an intensive care unit (ICU) stay [1–3]. The occurrence of systemic changes, including musculoskeletal, is well established in the literature due to prolonged immobility [4–6]. There is a reduction in tissue stiffness, cross-sectional area (CSA), and tendon thickness in disuse-induced situations [5–9]. Parry et al (2015) showed that ultrasound (US) detected changes in the quadriceps muscle correlate with strength and other health related losses observed in critically ill patients [10]. A recent study also demonstrated that only muscle area and thickness significantly decreased, without any modification in the quadriceps rectus femoris central tendon thickness in mechanically-ventilated patients [2]. Although assessment of the integrity of the tendon is fundamental in the critical care environment, to enable quadriceps actions, no studies have verified the reproducibility of measures taken by US image acquisition from the patellar and quadriceps tendons in patients admitted to the ICU.

When considering the morphological characteristics of the mentioned tendons in the ICU, imaging tests such as Magnetic Resonance Imaging (MRI) and diagnostic ultrasound are the preferred options as these techniques do not use ionizing radiation [11]. MRI is most commonly used for tendon assessment even if the choice represents an expensive and not very accessible exam [12]. For this reason, the use of high resolution US has also been indicated as an alternative exam [12,13].

Quadriceps tendon measures obtained by US have demonstrated similar accuracy to those obtained by MRI in adult patients with suspected quadriceps ruptures [12,14,15]. In addition, US measures can be recorded using portable devices, and are relatively inexpensive and easy to handle, a provident choice for assessing patients confined to bed [16,17].

The main technical limitation of US is related to the dependence on raters' perception which could lead to measurement errors and misinterpretation [18–24]. Tiny changes or displacements coming from the ultrasonic head placement, as well as different pressure or orientation of the transducer, may significantly influence image acquisition [24]. For example, a tendon image acquired slightly oblique to the longitudinal axis of the structure may appear thicker than the image taken on its true axis [19]. The mentioned features could lead to a premature conclusion that measurements taken from US are not reliable; however, this should be systematically verified.

Reproducibility refers to the extent to which repeated measurements provide similar results from different opinions (agreement) or repeated measures (reliability) [25]. Researchers have demonstrated acceptable reproducibility for US measures of thickness and CSA of the patellar tendon in ambulatory patients, expressing intraclass correlation coefficient (ICC) ranging from 0.70–0.95 and 0.68–0.99, respectively [19–22,26].

In addition, few reports have verified ultrasound imaging in measuring tendon for patients admitted to the ICU [2]. While there is high reproducibility of ultrasound measured quadriceps tendon thickness and CSA [19,27], there is a lack of literature evaluating the reproducibility of these measures in critically ill patients. Therefore, we aimed to determine intra- and inter-rater reproducibility of thickness, CSA and echogenicity measures of tendon sonographic images recorded from critically ill patients. Our hypothesis was that US image acquisition in patients with critical illnesses would be reproducible, regardless of the rater's experience.

Materials and methods

An observational blinded study with repeated measures was conducted to determine the intra- and inter-rater reproducibility of measures obtained by ultrasound images from patellar and

quadriceps tendons in sedated and critically ill patients admitted in the ICU of a Brazilian tertiary hospital. This study was approved by an institutional Ethics Committee of Fundação de Ensino e Pesquisa em Ciências da Saúde da Secretaria de Estado de Saúde do Distrito Federal (FEPECS / SES-DF) n° 1.768.479 in accordance with the Helsinki Declaration of 1975. Informed consent was signed by a close relative since all the patients were sedated.

Sample

We recruited adult patients (aged ≥ 18 years) who were critically ill, sedated, and submitted to mechanical ventilation. Patients with any kind of wound on the skin where the image acquisition took place or any known change in tendon morphology (e.g., ankylosing spondylitis or rheumatoid arthritis) were excluded [19]. A prior pilot study performed with 5 patients (results unpublished) recommended a sample size of 18 participants, considering a minimally acceptable coefficient of 0.700 for an expected coefficient of 0.900, supposing type I ($\alpha = 0.05$) and type II ($\beta = 0.20$) error rates [28]. Although 18 subjects were technically sufficient according to the power adopted (0.800), considering possible loss of data, we formed a convenience sample composed of 20 patients.

Raters. Two independent raters performed the techniques recommended for US image acquisition. Although both raters were physical therapists habituated to dealing with inpatient care, only one had experience in recording US measures, denominated as the experienced rater (KLC) for the purpose of this study. The novice rater (JC) received technical training to operate the US and acquire measures from the knee but did not have previous experience in this area. The practice consisted of a 20-minute session as described previously [29]. A basic technical explanation of the protocol and supervised performance in five patients was performed before data collection.

Study protocol

Image acquisition. The patellar and quadriceps tendons of the right knee of each eligible subject were studied. US images were acquired with a SonoSite M-Turbo portable ultrasound device (Sonosite, Inc., Bothell, WA, USA), equipped with a 2-dimensional, high-frequency linear array probe (HFL38, bandwidth: 13-6MHz, maximal scan depth: 6cm). Subjects were placed in the supine position with their knees in passive extension and neutral rotation 10 minutes before the onset of recording.

Reference marks were made on the skin, taking into account the total size of the quadriceps and patellar tendons to ensure minimal placement criteria. The array probes were placed 3 cm proximal to the superior pole of the patella for the quadriceps tendon [30] and at 25%, 50%, and 75% of the length of the patellar tendon [31]. The patellar tendon length was measured between the deep insertion in the patella and the deep insertion in the tibial tuberosity [31]. These landmarks were easily visible in the ultrasonographic image as hyperechogenic regions in bone insertion.

The raters positioned the probe perpendicularly and transversally (axial plane) over the marks, maintaining constant depth, gain, and ultrasound settings throughout the data collection period. Water-soluble transmission gel was used to decrease the pressure of the transducer on the skin [32] and optimize acoustic transmission [33]. We randomized the order of rater acquisition by employing a random numerical sequence generated on the website: <https://www.random.org>. Each rater acquired two sequential images on the same day, without checking the first one, as the quality of images was verified off-line.

The two images (first and second records) were made by each rater in order to investigate the intra-rater reliability, with the transducer decoupled from the skin and repositioned for the




Fig 1. Sonographic appearance of patellar and quadriceps tendons. Example of ultrasonography scans of the patellar (upper part) and quadriceps (lower part) tendons. The left panel depicts the different structures schematically. The dotted line represents the region of interest for the cross-sectional area and echogenicity of each tendon. Tendon thickness is symbolized by the dotted arrows. Tendon scans were conducted with a 4.9 cm depth. For example, considering the qualitative analysis performed in this study, A and C were qualified as “good” images; B and D were qualified as “bad” images.

<https://doi.org/10.1371/journal.pone.0219057.g001>

next record. The data were stored as files on the device itself and subsequently transferred to a computer for off-line processing. A third blinded researcher (PES) removed the stored files identifying the image. Subsequently, the file was evaluated by the experienced rater (KLC); sixteen images per individual were collected.

Analysis and experimental conditions. All images were analyzed with ImageJ software (National Institute of Health, Bethesda, MD, USA), allowing the raters to classified the images, as well as record the tendon thickness, CSA, and echogenicity for both the patellar and quadriceps tendons.

The experienced rater classified images as either: good–image with well-defined tendon borders–or bad–image with poorly-defined tendon borders. The opinions were employed in the inter-rater agreement analysis described below. The image quality analysis considered all landmarks; however, the average of the three measures obtained from the patellar tendon was calculated for the next analysis. The same image quality analysis was carried out at 2 different times, with a 7-day interval to verify the actual characterization of the image (Fig 1).

We calculated the thickness in three different locations along each image for both patellar and quadriceps tendons (Fig 1). The mean value was subsequently used for analyses. In turn, the CSA was measured by the trace technique consisting of the delimitation of the entire visible area of the tendon, excluding the peritendinous sheath (visible as a distinct, highly echogenic region both superficial and deep in the tendon) [19]. Echogenicity measurements were performed by the trace technique adopting the same area previously described for the calculation of CSA as the region of interest. Echogenicity was represented by a histogram on a gray scale with values ranging from 0 to 255 (0: black/no wave reflection; 255: white/total wave reflection).

For each outcome (thickness, CSA and echogenicity), first and second records were compared for each rater for patellar and quadriceps tendons, defining 6 experimental conditions for testing intra-rater reliability. For each variable, the analysis was carried out by the experienced and novice raters comparing single and double measures of the two tendons, defining a further 6 analyses to verify inter-rater reliability. The 6 experimental conditions of each analysis (intra and inter-rater) were divided by tendons (patellar and quadriceps) containing six Bland-Altman plots organized by pairs of repeated records in the columns (1st versus 2nd records for intra-rater or experienced versus novice measures for inter-rater) and by variables in the rows (thickness, CSA, and echogenicity). These measurements were performed twice consecutively by the experienced rater and the mean values considered for statistical analyses [34]. The intra-rater reliability was assessed between two subsequent repeated records taken by the raters (1st and 2nd records). For inter-rater analysis, the first and the average between the first and second repeated records defined the single and double measures, respectively.

Statistical analysis

The Shapiro-Wilk normality test identified non-Gaussian and Gaussian distribution for the data recorded from patellar and quadriceps tendons respectively. For this reason, the repeated measures recorded from the patellar tendon are represented as median followed by the lower and upper limits of the 95% Confidence Interval (non-parametric analysis) whereas mean and Standard Deviation (SD) are used for the data recorded from the quadriceps tendon (parametric analysis). The Wilcoxon matched-pairs signed test was used for all comparisons. Statistical significance was accepted at $p < 0.05$. Analysis was conducted using the first record (here called single) and the mean of the two records (double).

As mentioned, we considered two possibilities for the qualification of the image (good or bad), which were determined by the experienced examiner. The analysis of the agreement between experienced and novice records (inter-rater agreement) was performed using the Fleiss Kappa coefficient test applied for each region (proximal, middle, and distal regions) of the patellar tendon, and proximal region of the quadriceps tendon. We also calculated the gross agreement percentage (i.e., the percentage of observations in which the qualification of the records was the same). In accordance with Landis & Kock (1977), we considered values of $\kappa < 0$ as no agreement, $0 < \kappa \leq 0.20$ (poor agreement), $0.20 < \kappa \leq 0.40$ (fair agreement), $0.40 < \kappa \leq 0.60$ (moderate agreement), $0.60 < \kappa \leq 0.80$ (substantial agreement), and $0.80 \leq \kappa \leq 1.00$ (perfect agreement). A negative Kappa means that there is less agreement than would be expected by chance given the marginal distributions of ratings [35].

The intra and inter-rater reliabilities were processed for pairs of repeated measures; two images recorded by the same rater (1st and 2nd records) or two images recorded independently by different raters (experienced and novice). Initially, intraclass correlation coefficient (ICC) with a two-way random effects model either with single measure ($ICC_{2,1}$) or with average measures ($ICC_{2,2}$), calculated by taking an average of the two raters' measurements. Subsequently, the Bland-Altman (B&A) method was applied. The quality of the correlation detected by ICC was classified by the scale suggested by Lee et al (2012): 0 (absence), 0–0.19 (poor), 0.20–0.39 (weak), 0.40–0.59 (moderate), 0.60–0.79 (substantial), and ≥ 0.80 (almost complete) [36]. All statistical analyses were performed using SPSS for Mac (version 23, IBM, Chicago, Illinois, USA).

Results

Twenty critically ill patients (90% male) admitted to the ICU (diagnosis/condition: neurosurgical, medical, and trauma) constituted the convenience sample of this study. On the day of

Table 1. Measurement behaviour recorded in each of the 24 experimental conditions.

Experimental Conditions				Repeated Measures	
Raters	Tendons	Variables	Measures	Mean or Median	±SD or [min,max]
Experienced	Patellar	Thickness (cm)	single	0.36	[0.34,0.39]
			double	0.36	[0.33,0.39]
		CSA (cm ²)	single	1.24	[1.10,1.37]
			double	1.24	[1.12,1.38]
		Echogenicity (au)	single	55.39	[49.29,60.37]
			double	55.28	[49.43,61.51]
	Quadriceps	Thickness (cm)	single	0.90	± 0.16
			double	0.90	± 0.16
		CSA (cm ²)	single	2.68	± 0.54
			double	2.66	± 0.50
		Echogenicity (au)	single	52.68	± 12.95
			double	53.32	± 13.66
Novice	Patellar	Thickness (cm)	single	0.37	[0.33,0.41]
			double	0.37	[0.32,0.42]
		CSA (cm ²)	single	1.34	[1.14,1.46]
			double	1.30	[1.12,1.48]
		Echogenicity (au)	single	59.35	[47.32,61.84]
			double	58.64	[44.42,62.20]
	Quadriceps	Thickness (cm)	single	0.90	± 0.18
			double	0.90	± 0.16
		CSA (cm ²)	single	2.59	± 0.54
			double	2.59	± 0.54
		Echogenicity (au)	single	59.28	± 14.08
			double	58.86	± 14.09

Data recorded on the patellar tendon were obtained from twenty subjects ($n = 20$) whereas on the quadriceps tendon the sample was composed of sixteen ($n = 16$) due to the bad quality of the images. Data are presented as mean±SD (quadriceps tendon data) and median lower and upper limits of the 95% Confidence Interval (CI) of the median [min,max] (patellar tendon data). $P > 0.05$ for comparisons between experienced versus novice and single versus double for all reported variables.

Abbreviations: CSA—Cross-sectional area; SD—Standard Deviation; min—minimum; max—maximum.

<https://doi.org/10.1371/journal.pone.0219057.t001>

evaluation, subjects had a median age of 48 (35–55) years, body mass index of 25 (23–28) kg/m², median length of ICU stay of 11 (6–18) days, and duration of mechanical ventilation of 8 (4–18) days. The results of measurements of the patellar and quadriceps tendons did not present any significant differences between the experienced and novice raters or between single and double measurements for any of the reported variables (Table 1). On average, the experienced rater spent less time than the novice rater to acquire the ultrasound images (2.5 min vs 7.0 min, respectively).

Inter-rater agreement

The inter-rater analysis (Table 2) demonstrated that the majority of endpoints showed at least substantial agreement. The weakest agreement ($k = 0.571$) and the lowest gross agreement percentage (85%) were found for the second record for the quadriceps tendon considering statistically significant results. It should be noted that Kappa statistics were not computed in the middle of the patellar tendon length. No statistics were possible because all images in this tendon region were classified as “good”. There was no statistically significant agreement in the first record of quadriceps tendon and second record of the patellar distal region.

Table 2. Agreement between the quality of records acquired by two independent raters.

Ultrasound Images			Experienced vs Novice		
Image	Tendons	Regions	Kappa	95% CI	% Agreement
1 st record	Patellar	Proximal	1.000	[1.000,1.000]	100.00%
		Middle	Kappa could not be computed*		
		Distal	0.643	[0.006,1.000]	95.00%
	Quadriceps	Proximal	0.306	[-1.139,0.750]	75.00%
2 nd record	Patellar	Proximal	1.000	[1.000,1.000]	100.00%
		Middle	Kappa could not be computed*		
		Distal	-0.053	[-0.125,0.020]	90.00%
	Quadriceps	Proximal	0.571	[1.142,1.000]	85.00%

The experienced rater expressed its opinion (good or bad image) on the first (1st) and second (2nd) records. The asterisk (*) indicates the condition when the kappa could not be computed as all records were classified as good. Abbreviation: *vs*—*versus*.

<https://doi.org/10.1371/journal.pone.0219057.t002>

Intra- and inter-rater reliability

The magnitude of the reliability (ICC) of tendon thickness, CSA and echogenicity was classified as “almost complete” no matter the rater. When considering inter-rater reliability, ICCs were classified as “almost complete” when two images were considered for analysis. However, patellar reliability revealed no statistically significant correlations using a single US image ($p > 0.05$). All measures of correlation are demonstrated together with Bland-Altman plots (Figs 2–5).

Discussion

The present study confirmed the initial hypothesis demonstrating that the use of a standardized technique by experienced or novice raters resulted in reproducible measurements of tendon thickness, CSA, and echogenicity. Tendon imaging acquisition could be assessed by care providers with a minimal level of expertise in US devices. Thus, it may be possible to screen the effects of rehabilitative treatments on tendons.

The evaluation of US image quality through the kappa coefficient allowed us to establish the dimension of agreement beyond that expected by chance [37]. Agreement by chance can be demonstrated as gross agreement (% agreement in Table 2). For the distal patellar, there was a coincidence in the classification of 90% of the images acquired by different raters. However, this fact did not suggest statistical significance for the purpose of agreement analysis (is not a true agreement). It was observed that the statistically significant agreement was classified as moderate, reaching values of perfect agreement ($\kappa = 1.0$).

The inter-rater agreement on image quality showed that the middle region of the patellar tendon was the easiest region to obtain good quality images. However, the tendon distal region did not show statistical significance. The tendon borders in the distal region were more difficult to visualize, probably because of the tibial tuberosity. Coupling the US transducer to a prominent and rigid region at the extremities of the bones would have impaired complete image capture of the entire structure, possibly causing the borders to be unclear. Interestingly, the lowest agreement data were found for the quadriceps tendon. Tendon borders are less evident in the US images compared to patellar images. This may cause inconsistency in the classification of the image as being of good quality. Recently, the delimitation of the quadriceps tendon image was confirmed in the ultrasound image [30]. Despite being a crucial tendon for




Fig 2. Patellar tendon—Intra-rater Reliability (1st vs 2nd records). Bland-Altman plots of the intra-rater comparison between first (1st) and second (2nd) repeated records of the thickness (A and B), cross-sectional area (CSA, C and D), and echogenicity (E and F) taken on the patellar tendon. The images taken by the experienced rater are in the left column (A, C, and E) and plotted by white circles, whereas those taken by the novice rater are in the right column (B, D, and F) and plotted by black circles. Besides the zero lines, the bias line and random error lines showing the 95% limits of agreement are represented by dashed lines.

<https://doi.org/10.1371/journal.pone.0219057.g002>

the knee extensor mechanism, this type of evaluation is still incipient, and improvement and familiarization with this method could result in greater consistency of image acquisition.

The ICC values of the patellar and quadriceps tendons for thickness, CSA and echogenicity reached a degree of correlation classified as "almost complete", even considering the novice rater. These data are in agreement with previous studies that analyzed the thickness and CSA by US in individuals with orthopedic lesions in the outpatient setting [19–22,26]. However, no studies have assessed tendon reliability in highly complex treatment ICU settings. Ekizos et al. [27] stated that US is not a reliable instrument for measuring the CSA of the patellar tendon of




Fig 3. Quadriceps tendon—Intra-rater Reliability (1st vs 2nd records). Bland-Altman plots of the intra-rater comparison between first (1st) and second (2nd) repeated records of the thickness (A and B), cross-sectional area (CSA, C and D), and echogenicity (E and F) taken on the quadriceps tendon. The images taken by the experienced rater are in the left column (A, C, and E) and plotted by white circles, whereas those taken by the novice rater are in the right column (B, D, and F) and plotted by black circles. Besides the zero lines, the bias line and random error lines showing the 95% limits of agreement are represented by dashed lines.

<https://doi.org/10.1371/journal.pone.0219057.g003>

healthy young individuals. This discrepancy with our results can be justified by the different methods used for image acquisition. That study used the origin and insertion of the tendon (stable and easily visible points on the US), as well as the median (midway between the origin and insertion) and compared patellar tendon regions separately. In our study, the CSA was the average of three well-defined regions (25%, 50%, and 75% of tendon length) as previously proposed [6,31,38].

The Bland-Altman analysis showed that both repeated measures by the raters (intra-rater) and the repeated single and double measures between the raters (inter-rater) presented a

Inter-rater Reliability (experienced vs novice)




Fig 4. Patellar tendon-Inter-rater Reliability (experienced vs novice). Bland-Altman plots of the inter-rater comparison between repeated measures taken by experienced and novice raters, of the thickness (A and B), cross-sectional area (CSA, C and D), and echogenicity (E and F) on the patellar tendon. The single and double measures are respectively in the left (A, C, and E) and (B, D and F) right columns and plotted by white and black circles. Besides the zero lines, the bias line and random error lines showing the 95% limits of agreement are represented by dashed lines.

<https://doi.org/10.1371/journal.pone.0219057.g004>

predominant pattern of random variability. As the measurements were highly associated and limits of agreement (LOA) were small, the error was insignificant for most variables in both the patellar tendon and quadriceps tendon. This small error in repeated measures was described as a random type error associated with the A2 model of the B & A plot proposed by David and Giavarina [39]. In addition, there is no specific reason that can lead to this type of error.

Interestingly, intra-rater reliability seems to have a constant error in 3 groups of patients (5, 3, and 2 patients) and no error in 5 patients. The cases aligned in Fig 2B was named as the patient group. However, as the measures showed very high correlation (ICC 0.97), it is possible

Inter-rater Reliability (experienced vs novice)




Fig 5. Quadriceps tendon—Inter-rater Reliability (experienced vs novice). Bland-Altman plots of the inter-rater comparison between repeated images taken by experienced and novice raters, of the thickness (A and B), cross-sectional area (CSA, C and D), and echogenicity (E and F) on the quadriceps tendon. The single and double measures are respectively in the left (A, C, and E) and (B, D, and F) right columns and plotted by white and black circles. Besides the zero line, the bias line and random error lines forming the 95% limits of agreement are presented by dashed lines.

<https://doi.org/10.1371/journal.pone.0219057.g005>

to suggest that there is no specific reason to induce this type of error, i.e., these results are associated with a random error. The quadriceps tendon was also an example of random error with almost complete ICC (Fig 3 - graphs A and B). It seems that the novice examiner chose the visual references for thickness and did not change this parameter due to the smaller LOA amplitude, in contrast to the experienced rater.

Unexpectedly, slightly lower ICC values were found for the experienced rater compared to the novice considering quadriceps tendon. The average time of examination by the novice rater was practically 3 times longer than that of the experienced rater (7min vs. 2.5min), and it is possible to suggest that the longer time spent on image acquisition the better the image quality,

implying more reliable quantification of the structure. In contrast, it was not surprising not found statistically significance in some single inter-rater condition. This could be due to the bias of inherent variability in different raters, which has been described previously [38,39].

It is important to note that there was no statistically significant when only one measure was used for some measurements (Figs 4 and 5). However, thickness, CSA, and echogenicity raw values did not show any significant differences. The greatest difference between absolute inter-examiner values was approximately 10% for the echogenicity measure (Table 1). Probably, this small difference did not reflect any significant clinical repercussion. In fact, echogenicity represents the tendon by a gray scale where black color is no wave reflection and white color refers to total wave reflection. Anisotropy is an artefact and occurs when organized fibrils may reflect a majority of the insonating sound beam in a direction away from the transducer. This will cause the tendon to change from brightly hyperechoic to darkly hypoechoic [12]. To prevent this unwanted effect, the ultrasound probe should be positioned perpendicular to the structure being imaged [13]. Since the images were blindly acquired by two raters, it is likely that the angulation of the transducer used in the evaluation was not exactly the same. Thus, there may be some discrepancy in the results of echogenicity. However, it is emphasized that a high correlation was found for echogenicity probably because it was used a well-defined and standardized acquisition protocol. This find is in agreement with another study performed on muscle [10].

To our knowledge, this is the first study on US reliability of the quadriceps tendon in critically ill patients. We found studies that confirmed the validity of the instrument for the diagnosis of traumatic injury [40,41], and established the dimensions of the tendon in immature patients [34], as well as a review study on the sonographic image of the quadriceps tendon [12]. The instrument's ability to reproduce data should attest the use of this tool in both scientific and clinical research. Therefore, this study is the initial step to disseminating the use of US for serial evaluations of the quadriceps tendon in an intensive care setting.

Study limitations

Further studies should develop strategies or devices to better understand the uniqueness of the US assessment. For example, reliability of the marking procedure itself (e.g., joint position) and determining image settings (e.g., depth/gain, transducer type) are still potential unanticipated sources of error at the bedside. Besides, it should be noted that these results are applicable in the context of detecting longitudinal changes (where the same landmark is used over time). Further studies are required to determine the reproducibility of single time cross-sectional analysis, which involves the placement of a new landmark.

Conclusions

Ultrasound imaging acquisition measures taken by independent raters for tendon thickness, CSA, and echogenicity monitoring of critically ill patients are acceptable and not influenced by rater experience. It may be important to take time for image acquisition to increase reproducibility.

Acknowledgments

The authors thank Marco Aurélio Vaz (Neuromuscular Plasticity Research Group of the ESE-FID-UFRGS) for the critical revision regarding ultrasound assessment. We would like to thank the Institute of strategic health management of the Federal District ICU team.

Author Contributions

Conceptualization: Joana Castro, Paulo Eugênio Silva, Nicolas Babault, Rita de Cássia Marqueti, João Luiz Quagliotti Durigan.

Data curation: Joana Castro, Karina Livino de Carvalho, Paulo Eugênio Silva.

Formal analysis: Joana Castro, Karina Livino de Carvalho, Paulo Eugênio Silva.

Funding acquisition: João Luiz Quagliotti Durigan.

Investigation: Joana Castro, Rita de Cássia Marqueti, João Luiz Quagliotti Durigan.

Methodology: Joana Castro, Karina Livino de Carvalho, Paulo Eugênio Silva.

Project administration: Joana Castro, João Luiz Quagliotti Durigan.

Software: Emerson Fachin-Martins.

Supervision: Emerson Fachin-Martins, Nicolas Babault, Rita de Cássia Marqueti, João Luiz Quagliotti Durigan.

Writing – original draft: Joana Castro.

Writing – review & editing: Joana Castro, Emerson Fachin-Martins, Nicolas Babault, Rita de Cássia Marqueti, João Luiz Quagliotti Durigan.

References

1. Katari Y, Srinivasan R, Arvind P HS. Point-of-Care Ultrasound to Evaluate Thickness of Rectus Femoris, Vastus Intermedius Muscle, and Fat as an Indicator of Muscle and Fat Wasting in Critically Ill Patients in a Multidisciplinary Intensive Care Unit. *Indian J Crit Care Med*. 2018; 22(11):781–8. https://doi.org/10.4103/ijccm.IJCCM_394_18 PMID: 30598564
2. Hernández-Socorro CR, Saavedra P, López-Fernández JC R-SS. Assessment of Muscle Wasting in Long-Stay ICU Patients Using a New Ultrasound Protocol. *Nutrients*. 2018; 10(12).
3. Silva PE, Maldaner V, Vieira L, de Carvalho KL, Gomes H, Melo P, et al. Neuromuscular electrophysiological disorders and muscle atrophy in mechanically-ventilated traumatic brain injury patients: New insights from a prospective observational study. *J Crit Care*. 2018; 44:87–94. <https://doi.org/10.1016/j.jcrc.2017.10.026> PMID: 29078131
4. Zhang P, Hamamura K, Yokota H. A Brief Review of Bone Adaptation to Unloading. Vol. 6, *Genomics, Proteomics and Bioinformatics*. 2008. p. 4–7. [https://doi.org/10.1016/S1672-0229\(08\)60016-9](https://doi.org/10.1016/S1672-0229(08)60016-9) PMID: 18558381
5. Reeves ND, Maganaris CN, Ferretti G, Narici M V. Influence of 90-day simulated microgravity on human tendon mechanical properties and the effect of resistive countermeasures. *J Appl Physiol*. 2005; 98(6):2278–86. <https://doi.org/10.1152/jappphysiol.01266.2004> PMID: 15705722
6. Maganaris CN, Reeves ND, Rittweger J, Sargeant AJ, Jones DA, Gerrits K, et al. Adaptive response of human tendon to paralysis. *Muscle and Nerve*. 2006; 33(1):85–92. <https://doi.org/10.1002/mus.20441> PMID: 16216011
7. Reeves ND. Adaptation of the tendon to mechanical usage. *J Musculoskelet Neuronal Interact*. 2006; 6(2):174–80. PMID: 16849829
8. Carroll CC, Dickinson JM, Haus JM, Lee G a, Hollon CJ, Aagaard P, et al. Influence of aging on the in vivo properties of human patellar tendon. *J Appl Physiol*. 2008; 105(6):1907–15. <https://doi.org/10.1152/jappphysiol.00059.2008> PMID: 18927271
9. Heinemeier KM, Kjaer M. In vivo investigation of tendon responses to mechanical loading. Vol. 11, *Journal of Musculoskeletal Neuronal Interactions*. 2011. p. 115–23. PMID: 21625048
10. Parry SM, El-Ansary D, Cartwright MS, Sarwal A, Berney S, Koopman R, et al. Ultrasonography in the intensive care setting can be used to detect changes in the quality and quantity of muscle and is related to muscle strength and function. *J Crit Care*. 2015; 30(5):1151.e9–1151.e14.
11. Weinreb JH, Sheth C, Apostolakis J, McCarthy M-B, Barden B, Cote MP, et al. Tendon structure, disease, and imaging. *Muscles Ligaments Tendons J* [Internet]. 2014; 4(1):66–73. Available from: <http://www.pubmedcentral.nih.gov/articlerender.fcgi?artid=4049653&tool=pmcentrez&rendertype=abstract> PMID: 24932450

12. Alves TI, Girish G, Kalume Brigido M, Jacobson JA. US of the Knee: Scanning Techniques, Pitfalls, and Pathologic Conditions. *RadioGraphics* [Internet]. 2016; 36(6):1759–75. Available from: <https://doi.org/10.1148/rg.2016160019> PMID: 27726755
13. De Maeseneer M, Marcelis S, Boulet C, Kichouh M, Shahabpour M, De Mey J, et al. Ultrasound of the knee with emphasis on the detailed anatomy of anterior, medial, and lateral structures. Vol. 43, *Skeletal Radiology*. 2014. p. 1025–39. <https://doi.org/10.1007/s00256-014-1841-6> PMID: 24622925
14. Henderson REA, Walker BF, Young KJ. The accuracy of diagnostic ultrasound imaging for musculoskeletal soft tissue pathology of the extremities: a comprehensive review of the literature. *Chiropr Man Therap* [Internet]. 2015; 23:31. Available from: <http://www.pubmedcentral.nih.gov/articlerender.fcgi?artid=4634582&tool=pmcentrez&rendertype=abstract> <https://doi.org/10.1186/s12998-015-0076-5> PMID: 26543553
15. Bianchi S, Zwass A, Abdelwahab IF, Banderali A. Diagnosis of tears of the quadriceps tendon of the knee: Value of sonography. *Am J Roentgenol*. 1994; 162(5):1137–40.
16. Mourtzakis M, Wischmeyer P. Bedside ultrasound measurement of skeletal muscle. *Curr Opin Clin Nutr Metab Care* [Internet]. 2014; 17(5):389–95. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/25023190> <https://doi.org/10.1097/MCO.0000000000000088> PMID: 25023190
17. Chauvin NA, Ho-Fung V, Jaramillo D, Edgar JC, Weiss PF. Ultrasound of the joints and entheses in healthy children. *Pediatr Radiol*. 2015 Aug; 45(9):1344–54. <https://doi.org/10.1007/s00247-015-3313-0> PMID: 25744571
18. O'Connor PJ, Grainger AJ, Morgan SR, Smith KL, Waterton JC, Nash AF. Ultrasound assessment of tendons in asymptomatic volunteers: A study of reproducibility. *Eur Radiol*. 2004; 14(11):1968–73. <https://doi.org/10.1007/s00330-004-2448-4> PMID: 15309497
19. Dudley-Javoroski S, McMullen T, Borgwardt MR, Peranich LM, Shields RK. Reliability and Responsiveness of Musculoskeletal Ultrasound in Subjects with and without Spinal Cord Injury. *Ultrasound Med Biol*. 2010; 36(10):1594–607. <https://doi.org/10.1016/j.ultrasmedbio.2010.07.019> PMID: 20800961
20. Gellhorn AC, Carlson MJ. Inter-Rater, Intra-Rater, and Inter-Machine Reliability of Quantitative Ultrasound Measurements of the Patellar Tendon. *Ultrasound Med Biol*. 2013; 39(5):791–6. <https://doi.org/10.1016/j.ultrasmedbio.2012.12.001> PMID: 23465140
21. Özçakar L, Kara M, Tekin L, Karanfil Y, Esen E, Utku B, et al. Effect of supervision on ultrasonographic measurements. A blinded randomized cross-over study. *Eur J Phys Rehabil Med*. 2013;
22. Skou ST, Aalkjaer JM. Ultrasonographic measurement of patellar tendon thickness—a study of intra- and interobserver reliability. *Clin Imaging*. 2013; 37(5):934–7. <https://doi.org/10.1016/j.clinimag.2013.01.007> PMID: 23759209
23. Mc Auliffe S, Mc Creesh K, Purtill H, O'Sullivan K. A systematic review of the reliability of diagnostic ultrasound imaging in measuring tendon size: Is the error clinically acceptable? *Physical Therapy in Sport*. 2016;
24. Vieira A, Siqueira AF, Ferreira-Junior JB, Pereira P, Wagner D, Bottaro M. Ultrasound imaging in women's arm flexor muscles: intra-rater reliability of muscle thickness and echo intensity. *Braz J Phys Ther* [Internet]. 2016; Available from: <http://dx.doi.org/10.1590/bjpt-rbf.2014.0186>
25. Mokkink LB, Terwee CB, Patrick DL, Alonso J, Stratford PW, Knol DL, et al. The COSMIN checklist for assessing the methodological quality of studies on measurement properties of health status measurement instruments: An international Delphi study. *Qual Life Res*. 2010; 19(4):539–49. <https://doi.org/10.1007/s11136-010-9606-8> PMID: 20169472
26. Toprak U, Üstüner E, Uyanik S, Aktaş G, Kinikli Gİ, Baltacı G, et al. Comparison of ultrasonographic patellar tendon evaluation methods in elite junior female volleyball players: Thickness versus cross-sectional area. *Diagnostic Interv Radiol*. 2012; 18(2):200–7.
27. Ekizos A, Papatzika F, Charcharis G, Böhm S, Mersmann F, Arampatzis A. Ultrasound does not provide reliable results for the measurement of the patellar tendon cross sectional area. *J Electromyogr Kinesiol*. 2013; 23(6):1278–82. <https://doi.org/10.1016/j.jelekin.2013.08.004> PMID: 24021864
28. Walter SD, Eliasziw M, Donner A. Sample size and optimal designs for reliability studies. *Stat Med*. 1998;
29. Zaidman CM, Wu JS, Wilder S, Darras BT, Rutkove SB. Minimal training is required to reliably perform quantitative ultrasound of muscle. *Muscle and Nerve*. 2014; 50(1):124–8. <https://doi.org/10.1002/mus.24117> PMID: 24218288
30. de Carvalho KL, Silva PE, Castro J, Babault N, Durigan JLQ, de Cássia Marqueti R. Height, Weight, and Age Predict Quadriceps Tendon Length and Thickness in Skeletally Immature Patients: Letter to the Editor. *Am J Sports Med* [Internet]. 2017; 45(9):NP26–NP26. Available from: <https://doi.org/10.1177/0363546517717699> PMID: 28714794

31. O'Brien TD, Reeves ND, Baltzopoulos V, Jones D a, Maganaris CN. Mechanical properties of the patellar tendon in adults and children. *J Biomech* [Internet]. 2010; 43(6):1190–5. Available from: <http://www.ncbi.nlm.nih.gov/pubmed/20045111> <https://doi.org/10.1016/j.jbiomech.2009.11.028> PMID: 20045111
32. Arts IMP, Pillen S, Schelhaas HJ, Overeem S, Zwarts MJ. Normal values for quantitative muscle ultrasonography in adults. *Muscle and Nerve*. 2010; 41(1):32–41. <https://doi.org/10.1002/mus.21458> PMID: 19722256
33. Kanehisa H, Ito M, Kawakami Y, Fukunaga T, Miyatani M. The accuracy of volume estimates using ultrasound muscle thickness measurements in different muscle groups. *Eur J Appl Physiol*. 2004 Mar; 91(2–3):264–72. <https://doi.org/10.1007/s00421-003-0974-4> PMID: 14569399
34. Todd DC, Ghasem AD, Xerogeanes JW. Height, weight, and age predict quadriceps tendon length and thickness in skeletally immature patients. *Am J Sports Med*. 2015; 43(4):945–52. <https://doi.org/10.1177/0363546515570620> PMID: 25721535
35. Landis JR, Koch GG. The Measurement of Observer Agreement for Categorical Data. *Biometrics* [Internet]. 1977; 33(1):159. Available from: <http://www.jstor.org/stable/2529310?origin=crossref> PMID: 843571
36. Lee KM, Lee J, Chung CY, Ahn S, Sung KH, Kim TW, et al. Pitfalls and Important Issues in Testing Reliability Using Intraclass Correlation Coefficients in Orthopaedic Research. *Clin Orthop Surg*. 2012; 4(2):149. <https://doi.org/10.4055/cios.2012.4.2.149> PMID: 22662301
37. Sim J, Wright CC. The Kappa Statistic in Reliability Studies: Use, Interpretation, and Sample Size Requirements. *Phys Ther* [Internet]. 2005; 85(3):257–68. Available from: <http://www.ptjournal.org/cgi/content/abstract/85/3/257%5Cnhttp://www.ptjournal.org/cgi/content/full/85/3/257> PMID: 15733050
38. Onambele GN, Burgess K, Pearson SJ. Gender-specific in vivo measurement of the structural and mechanical properties of the human patellar tendon. *J Orthop Res*. 2007; 25(12):1635–42. <https://doi.org/10.1002/jor.20404> PMID: 17568426
39. Atkinson G, Nevill AM. Statistical methods for assessing measurement error (reliability) in variables relevant to sports medicine. *Sport Med*. 1998; 26(4):217–38.
40. La S, Fessell DP, Femino JE, Jacobson J a, Jamadar D, Hayes C. Sonography of partial-thickness quadriceps tendon tears with surgical correlation. *J Ultrasound Med*. 2003; 22(12):1323–1329-1331. PMID: 14682419
41. Swamy GN, Nanjayan SK, Yallappa S, Bishnoi A, Pickering SAW. Is ultrasound diagnosis reliable in acute extensor tendon injuries of the knee? Vol. 78, *Acta Orthopaedica Belgica*. 2012. p. 764–70. PMID: 23409573